

Numar curent: 13

Aparat	Frecventmetru electronic#2.03.03	Motiv ver/et	Periodica
Model	FE21	Metoda ver/et	Comparare directa
Serie	82512	Norma ver/et	N.T.M 2-05-1990
Anul fabricatiei	1980	Valoare verificare	56
Domeniu	0-1000Hz	Verificat extern	Nu
Clasa precizie	2	Etalon	Nu
Incadrare	Timp-Frecventa	Etalonat cu	0
Departament	Depart. Calit.	Verificator	INM Buc.
Utilizator	Balasa Dtru.	Reparator	Leonte Maria
Interval ver/et	24 luni	Stare aparat	Functional
Interval ver/et LO	2 ani	Firma constructoare	Japonia
Verificat la	28.02.2005	Reparat la	15.02.2007
Se verifica la	28.02.2007	Defect, reparatie	Carcasa sparta, inlocuit carcasa
Rezultat verificare	Corespunde	Data inregistrarii	23.04.1980
Numar document	CE 063	Client	Fara
		Observatii	